

Czech Metrology Institute Notified Body No. 1383

V 3112

Okružní 31, 638 00 Brno, Czech Republic tel. +420 545 555 111, fax +420 545 222 728

EU-TYPE EXAMINATION CERTIFICATE

Number: TCM 142/13 - 5009

Addition 3

This addition replaces all previous versions of this certificate in full wording.

Page 1 from 11 pages

In accordance:

with Directive 2014/32/EU of the European Parliament and of the Council on the harmonisation of the laws of the Member States relating to the making available on the market of measuring instruments (implemented in Czech Republic by Government Order No. 120/2016 Coll.).

Manufacturer:

ECOMESS Sp. z o.o. ul. Szczawińska 42c 95-100 Zgierz Poland

For:

water meter - single jet, dry and electronic dial

Type: Picoflux and i Flux

Accuracy class: 2

Temperature class: T30, T50, T70, T90 or T30/90

Valid until:

27 January 2023

Document No:

0511-CS-A004-13

Description:

Essential characteristics, approved conditions and special conditions, if any, are described in

this certificate.

Date of issue:

21 August 2018

Certificate approved by:

Tolingion Matthe

RNDr. Pavel Klenovský